

# 130 - 4020 Viking Way  
Richmond, BC V6V 2L4 CANADA

We,

**Datrend Systems Inc,**

declare in exclusive responsibility that the product:

Model: **vPad-A1 Modular Patient Simulator<sup>1</sup>**

is in conformity with the essential requirements and other relevant requirements of:

**Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

**Electromagnetic Compatibility Regulations 2016**

**Electrical Equipment (Safety) Regulations 2016**

Designated standards:

**EN 61326-1:2013** *Electrical equipment for measurement, control and laboratory use - EMC requirements*  
**EN 61010-1:2010** *Safety requirements for electrical equipment for measurement, control and laboratory use*  
**EN 63000:2018** *Technical documentation for assessment of electrical and electronic products with respect to restriction of hazardous substances*

Signed for, and on behalf of, Datrend Systems Inc.



Keith Ching  
Director of Manufacturing Technology

Date of Original Declaration: 2022-Jul-21

Date of Declaration: 2025-May-1

Place of Issue: Richmond BC Canada

---

<sup>1</sup> Comprising in combination the A1 Base Module and any one or more of the following simulation modules: vPad-PS Patient Simulator; vPad-O2 Pulse Oximeter Tester; and/or vPad-NIBP Non-Invasive Blood Pressure Simulator