

# 130 - 4020 Viking Way  
Richmond, BC V6V 2L4 CANADA

We,

**Datrend Systems Inc,**

declare in exclusive responsibility that the product:

**Model: vPad-CO2 End Tidal Carbon Dioxide (EtCO2) Breath Simulator**

is in conformity with the essential requirements and other relevant requirements of:

**Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012**

**Electromagnetic Compatibility Regulations 2016**

**Electrical Equipment (Safety) Regulations 2016**

Designated standards:

**EN 61326-1:2013** *Electrical equipment for measurement, control and laboratory use - EMC requirements*  
**EN 61010-1:2010** *Safety requirements for electrical equipment for measurement, control and laboratory use*  
**EN 63000:2018** *Technical documentation for assessment of electrical and electronic products with respect to restriction of hazardous substances*

Signed for, and on behalf of, Datrend Systems Inc.



Lloyd Van't Haaff  
Quality Manager

Date of Declaration: 2023-Jul-27

Place of Issue: Richmond BC Canada