

EC Declaration of Conformity
EG Konformitäts-Erklärung
Déclaration CE de Conformité
Declaración de Conformidad tipo EC

We, Datrend Systems Inc, declare in exclusive responsibility that the product:
Wir, Datrend Systems Inc, erklären in alleiniger Verantwortung, daß das Produkt:
Nous, Datrend Systems Inc, déclarons sous notre seule responsabilité que le produit:
Datrend Systems Inc declaramos en la responsabilidad exclusiva que el producto:

Model / Modell / Modèle / Modelo: **vPad-A1 Modular Patient Simulator¹**

to which this declaration relates is in conformity with the following Directives:
auf die sich diese Erklärung bezieht, steht im Einklang mit den folgenden Richtlinien:
auquel se réfère cette déclaration est conforme aux Directives suivantes:
a cuál se relaciona este declaración es conforme a las siguientes Directivas:

2014/30/EU - Electromagnetic Compatibility (EMC)

2014/35/EU - Low Voltage Directive (LVD)

2011/65/EU - RoHS Directive

2023/1542/EU - Battery Regulation²

Harmonized standards used:
Verwendete harmonisierte Standards:
Normes harmonisées utilisées:
Normas armonizados utilizados:

EN 61326-1:2013 *Electrical equipment for measurement, control and laboratory use - EMC requirements*

EN 61010-1:2010 *Safety requirements for electrical equipment for measurement, control and laboratory use*

EN 63000:2018 *Technical documentation for assessment of electrical and electronic products with respect to restriction of hazardous substances*

Signed for, and on behalf of, Datrend Systems Inc.



Keith Ching
Director of Manufacturing Technology

EU Authorised Representative:
Comply Express Unipessoal Limitada,
StartUp Madeira, EV141, Campus da Penteada, 9020 105
Funchal, Portugal

Date of Original Declaration: 2016-Aug-22

Date of Declaration: 2026-Jun-26

Place of Issue: Richmond BC Canada

¹ Comprising in combination with the vPad-A1 Base Module and any one or more of the following modules: vPad-PS Patient Simulator; vPad-O2 Pulse Oximeter Tester; and/or vPad-NIBP Non-Invasive Blood Pressure Simulator.

² Only the vPad-A1 Base Module and vPad-NIBP Non-Invasive Blood Pressure Simulator contain portable battery. Both modules conform to applicable Article 6, 9, 11, and 13 related to portable batteries.