▼ vPad-O2™ - Performance Specifications

Saturation (SpO₂):

Range: 30% to 100% Increments: 1%

Presets: 6, user definable

Range of adjustment and presets may vary according to pulse oximeter specifications

SpO₂ Accuracy:

Saturation within DUT specified range: ±1 count + specified accuracy of the DUT

Heart Rate:

Range: 20 to 300 BPM Increments: 1 BPM Presets: 6, user definable

Accuracy: ± 0.25 BPM (sync mode)

otherwise, ±1 BPM

Pulse Amplitude:

Range: 0 to 100% Increments: 1% steps. Presets: 6, user definable Accuracy: ± 1%

Signal Artifact:

Four preset simulations:

Movement

Tapping (Spike artifact) Shivering (Tremor artifact) Shake Table (2.5Hz Sinewave

Auto Presets:

Unlimited preset patient simulations

Default Auto Presets:

Normal Adult

Hypoxia

Movement Artifact

Tachycardia

Bardycardia

Neonate

Low Perfusion

No perfusion

Tremor (Shivering Artifact)

Alarm Tests:

Automated test sequences for determining oximeter alarm response time to:

Low Saturation Low Heart Rate High Heart Rate Low Perfusion Signal Artifact

Five defaults, plus unlimited programmable

alarm sequences

Communication / User Interface:

via vPad-A1 Base Unit Android 5" tablet:

> Touchscreen User Interface Wired (USB) or Bluetooth mode

WiFi

16 GB memory

Dual XBUS for Datrend test automation

Power Supply:

via vPad-A1 Base Unit External AC adapter

Internal rechargeable Li-Ion batteries (for 10 hrs of

simulation with full charge)

Dimensions:

98mm x 208mm x 30mm (3.85" x 8.2" x 1.18") A1 Base 90mm x 160mm x 24mm (3.54" x 6.3" x 0.95") SpO₂ Probe

Weight:

440g (0.96lb) A1 base 122g (0.27lb) SpO₂ Probe

Environment:

15°C to 40°C, 10% to 90% RH, Indoor Use Only, Category II

All specifications subject to change without notice.



130 - 4020 Viking Way, Richmond BC, Canada V6V 2L4

Phone: 604.291.7747 • Fax: 604.294.2355 Toll-free (North America only): 800.667.6557

Email: customerservice@datrend.com © DATREND SYSTEMS INC. SEPTEMBER 2016